



EMC EMISSIONS - TEST REPORT (Full)

Test Report No. **3152524DEN-002** Issue Date: Tuesday 27/May/2008

Model / Serial No. CP-100-24-N / SN: A002

Product Type X-Ray Collimator

Client Collimare

Manufacturer Collimare

License holder Collimare

Address 13406 W. 60th Place
Arvada, CO 80004

Test Criteria Applied **CISPR 11: 2004 Class A**
 Test Result **PASS** SPECIFICATION FOR LIMITS AND METHODS OF MEASUREMENT OF RADIO DISTURBANCE CHARACTERISTICS OF INDUSTRIAL, SCIENTIFIC AND MEDICAL (ISM) RADIO-FREQUENCY EQUIPMENT
 Test Project Number **3152524** Report Includes: IEC/EN 61000-3-2 and IEC/EN 61000-3-3 Where applicable
 References
 Total Pages
 Including **18**
 Appendices:


 Tested By : Ty Orosco


 Reviewed By : Michael Spataro

REVISION SUMMARY - The following changes have been made to this Report:

Rev.	Revision Statement	Author	Revision Date
	Initial Release of Document	See above	See above

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to copy or distribute this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.



Lab Number:
SL2-IN-E-029R

Lab Code: 200264-0

The entity logos above are for reference only and may not apply to this test report.